

Certificate of Analysis

Name: Protein Hydrolysate Standard, 2.5 mM
Item: 80-6002-67-1
Lot: 0006714248

Lot Issue Date: 06-Dec 2022
Expiration Date: 31-Dec 2025

This Reference Material was manufactured and verified on behalf of Biochrom Ltd by a Registered Quality Control laboratory in accordance with the ISO Standards listed below.

The true value and the uncertainty value at the 95% confidence level for each analyte are listed in the table. These were calculated from the gravimetric measurements made during the manufacturing process. All Balances were calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001. Calibrated Class A glassware was used in the manufacturing of these standards.

Analyte	CAS#	True Value	True Value
Taurine	000107-35-7	0.313 ± 0.002 g/L	2.50 mM
L-Aspartic acid	000056-84-8	0.334 ± 0.002 g/L	2.51 mM
L-Threonine	000072-19-5	0.299 ± 0.001 g/L	2.51 mM
L-Serine	000056-45-1	0.263 ± 0.001 g/L	2.50 mM
L-Glutamic acid	000056-86-0	0.369 ± 0.002 g/L	2.51 mM
L-Proline	000147-85-3	0.289 ± 0.001 g/L	2.51 mM
Glycine	000056-40-6	0.188 ± 0.001 g/L	2.50 mM
L-Alanine	000056-41-7	0.223 ± 0.001 g/L	2.50 mM
L-Valine	000072-18-4	0.293 ± 0.001 g/L	2.50 mM
L-Cystine	000056-89-3	0.301 ± 0.002 g/L	1.25 mM
L-Methionine	000063-68-3	0.373 ± 0.002 g/L	2.50 mM
L-Isoleucine	000073-32-5	0.328 ± 0.002 g/L	2.50 mM
L-Leucine	000061-90-5	0.329 ± 0.002 g/L	2.51 mM
L-Tyrosine	000060-18-4	0.455 ± 0.002 g/L	2.51 mM
L-Phenylalanine	000063-91-2	0.414 ± 0.002 g/L	2.51 mM
ammonium chloride	012125-02-9	0.134 ± 0.001 g/L	2.50 mM
L-Lysine	000056-87-1	0.365 ± 0.002 g/L	2.50 mM
L-Histidine	000071-00-1	0.389 ± 0.002 g/L	2.51 mM
L-Arginine	000074-79-3	0.438 ± 0.002 g/L	2.51 mM

Matrix: 0.1 N Hydrochloric Acid

(continued)

Name: Protein Hydrolysate Standard, 2.5 mM
Catalog No: 80-6002-67-1
Lot No: 0006714248

Lot Issue Date: 06-Dec 2022
Expiration Date: 31-Dec 2025

Quality Certifications.

This Reference Material was prepared by Agilent Technologies Inc., RI, USA, in compliance with the following ISO Standards:

- ISO 9001 Quality Management Systems Requirements (TUV USA Cert. No. 56 100 18560026)
- ISO 17034 General Requirements for the Competence of Reference Material Producers (ANAB Cert. No. AR-1936)
- ISO 17025 General Requirements for the Competence of Testing and Calibration Laboratories (ANAB Cert. No. AT-1937)



Monica Bourgeois
QMS Representative