

# CERTIFICATE OF ANALYSIS

For

## Lithium Loading Buffer NC

Part Number: 80-6003-07

Batch (Control) Number: P27758NC

Quantity: 2 Litres

Expiry Date: 02 July 2024

The above product gave the following results on analysis:

pH = 2.219 @ 20.0°C

pH Specification = 2.18 - 2.22 @ 20.0°C

Conductivity = 16.76 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 16.1 - 18.3 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:

David Thomas

Date of Issue: 13 July 2021

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.