

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 1

Part Number: 80-2038-15

Batch (Control) Number: P27781

Quantity: 2 Litres

Expiry Date: 01 December 2024

The above product gave the following results on analysis:

pH = 2.791 @ 20.0°C

pH Specification = 2.78 - 2.82 @ 20.0°C

Conductivity = 14.85 mScm⁻¹ @ 20.0°C

Conductivity Specification = 14.4 - 16.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 20 December 2021

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.