

# CERTIFICATE OF ANALYSIS

For

## EZ Nin Reagent

Part Number: 80-6000-12

Batch (Control) Number: R27811

Quantity: 2 Litres

Expiry Date: 21 July 2025

The above product gave the following results on a standard Lithium High-Performance System:

Baseline Drift:

-0.07  $\Delta\text{mV}/\Delta\text{t}$  at 440nm, 0.01  $\Delta\text{mV}/\Delta\text{t}$  at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Chemical QC Signature:

Date of Issue: 26 July 2022

Neil Booth

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.