

CERTIFICATE OF ANALYSIS

For

Ultrsolve Plus NC

Part Number: 80-6003-09

Batch (Control) Number: R27821NC

Quantity: 1.75 Litres

Expiry Date: 10 October 2025

The above product gave the following results on analysis:

pH = 5.640 @ 20.0°C

pH Specification = 5.55 - 5.75 @ 20.0°C

Conductivity = 19.2 mScm⁻¹ @ 20.0°C

Conductivity Specification = 17.5 - 23.5 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

David Thomas

Date of Issue: 17 October 2022

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.