

## Certificate of Analysis

Name: Oxidised Hydrolysate Standard, 2.5 mM  
Item: 80-6002-68  
Lot No: 0006740252

Lot Issue Date: 30-May 2023  
Expiration Date: 31-May 2026

This Reference Material was manufactured on behalf of Biochrom Ltd. in accordance with the ISO Standards and certifications listed below.

The true value and the uncertainty value at the 95% confidence level for each analyte are listed in the table. These were calculated from the gravimetric measurements made during the manufacturing process. All Balances were calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 17025. Calibrated Class A glassware was used in the manufacturing of these standards.

Analyte	CAS#	True Value	True Value
Taurine	000107-35-7	0.313 ± 0.002 g/L	2.50 mM
L-Aspartic acid	000056-84-8	0.334 ± 0.002 g/L	2.51 mM
L-Threonine	000072-19-5	0.299 ± 0.001 g/L	2.51 mM
L-Serine	000056-45-1	0.263 ± 0.001 g/L	2.50 mM
L-Glutamic acid	000056-86-0	0.369 ± 0.002 g/L	2.51 mM
L-Proline	000147-85-3	0.289 ± 0.001 g/L	2.51 mM
Glycine	000056-40-6	0.189 ± 0.001 g/L	2.52 mM
L-Alanine	000056-41-7	0.223 ± 0.001 g/L	2.50 mM
L-Valine	000072-18-4	0.294 ± 0.001 g/L	2.51 mM
L-Cystine	000056-89-3	0.301 ± 0.002 g/L	1.25 mM
L-Methionine	000063-68-3	0.375 ± 0.002 g/L	2.51 mM
L-Isoleucine	000073-32-5	0.329 ± 0.002 g/L	2.51 mM
L-Leucine	000061-90-5	0.329 ± 0.002 g/L	2.51 mM
L-Tyrosine	000060-18-4	0.455 ± 0.002 g/L	2.51 mM
D,L-Methionine sulfoxide	062697-73-8	0.414 ± 0.002 g/L	2.51 mM
L-Methionine sulfone	007314-32-1	0.455 ± 0.002 g/L	2.51 mM
L-Cysteic acid monohydrate (as L-Cysteic acid)	023537-25-9	0.425 ± 0.002 g/L	2.51 mM
L-Phenylalanine	000063-91-2	0.414 ± 0.002 g/L	2.51 mM
ammonium chloride	012125-02-9	0.134 ± 0.001 g/L	2.50 mM
L-Ornithine monohydrochloride	003184-13-2	0.424 ± 0.002 g/L	2.51 mM
L-Lysine	000056-87-1	0.366 ± 0.002 g/L	2.50 mM
L-Histidine	000071-00-1	0.390 ± 0.002 g/L	2.51 mM
L-Arginine	000074-79-3	0.437 ± 0.002 g/L	2.51 mM

Matrix: 0.1 N Hydrochloric Acid

(continued)

Name: Oxidised Hydrolysate Standard, 2.5 mM  
Item: 80-6002-68  
Lot No: 0006740252

Lot Issue Date: 30-May 2023  
Expiration Date: 31-May 2026

Safety Data Sheets: Can be found at <https://www.biochrom.co.uk/msdssearch.asp>

Quality Certifications.

This Reference Material was prepared by Agilent Technologies Inc., RI, USA, in compliance with the following ISO Standards:

- ISO 9001:2015 Quality Management Systems Requirements (TUV USA Certificate No: 951215321)
- ISO 17025:2017 General Requirements for the Competence of Reference Material Producers (ANAB Cert. No. AR-1937)
- ISO 17034:2016 General Requirements for the Competence of Testing and Calibration Laboratories (ANAB Cert. No. AT-1936)



X  
Date 2 May 2024  
Signatory Name Nitya Shetty  
Signatory Title Vice President, Global R&D