

CERTIFICATE OF ANALYSIS

For

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Batch (Control) Number: U27883

Quantity: 250 mL

Expiry Date: 10 April 2027

The above product gave the following results on analysis:

Baseline Drift:

-0.021 $\Delta mV/\Delta t$ at 440nm, 0.047 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Colorimetric Test:

Absorption @ 485nm = 0.045 A

Absorption @ 600nm = 0.015 A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Date of Issue: 23 April 2024

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.