

CERTIFICATE OF ANALYSIS

For

Lithium Loading Buffer NC

Part Number: 80-6003-07

Batch (Control) Number: T27855NC

Quantity: 2 Litres

Expiry Date: 19 June 2026

The above product gave the following results on analysis:

pH = 2.208 @ 20.0°C

pH Specification = 2.18 - 2.22 @ 20.0°C

Conductivity = 16.53 mScm⁻¹ @ 20.0°C

Conductivity Specification = 16.1 - 18.3 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

David Thomas

Date of Issue: 20 June 2023

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.