

CERTIFICATE OF ANALYSIS

For

Sodium Regeneration Buffer 6

Part Number: 80-2037-57

Batch (Control) Number: W27914

Quantity: 1 Litre

Expiry Date: 21 February 2028

The above product gave the following results on analysis:

Conductivity = 72.3 mScm⁻¹ @ 20.0°C

Conductivity Specification = 68.5 - 75.7 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 10 March 2025

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
Web Site: http://www.biochrom.co.uk			

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.